















<p>A-09 Z-valine</p>  <p>OA-3200 Column:4.0mmI.D.×25cm Mobile phase:0.005mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p> 	<p>A-10 Z-norvaline</p>  <p>OA-3200 Column:4.0mmI.D.×25cm Mobile phase:0.005mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p> 		
<p>OA-3200 1.13 OA-4400 1.08 { L, D } OA-3300 1.07</p>	<p>OA-4500 1.06 { L, D } OA-4700 1.06</p>	<p>OA-3200 1.13 OA-3300 1.09</p>	
<p>A-11 Z-leucine</p>  <p>OA-3200 Column:4.6mmI.D.×25cm Mobile phase:0.01mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm Injection:5µL (5mg/mL)</p> 	<p>A-12 Z-serine</p>  <p>OA-3200 Column:4.0mmI.D.×25cm Mobile phase:0.01mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p> 		
<p>OA-3200 1.21 { D, L } OA-3300 1.07</p>		<p>OA-3200 1.09 OA-3100 1.07 OA-4400 1.06</p>	
<p>A-13 Z-asparagine</p>  <p>OA-3200 Column:4.0mmI.D.×25cm Mobile phase:0.005mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p> 	<p>A-14 Z-phenylalanine</p>  <p>OA-3300 Column:4.0mmI.D.×25cm Mobile phase:0.01mol/L ammonium acetate in methanol Flow rate:0.5mL/min Detector:UV254nm</p> 		
<p>OA-3100 1.12 OA-3200 1.11</p>		<p>OA-3300 1.08 OA-3200 1.07 OA-4400 1.07 { L, D }</p>	<p>OA-4500 1.04</p>
<p>A-15 N-Dansylvaline</p>  <p>OA-7100 Column:4.6mmI.D.×25cm Mobile phase:30% acetonitrile in 0.1% TEAA pH5.5 Flow rate:0.7mL/min Detector:UV254nm Injection:2µL (2mg/mL)</p> 			
<p>OA-7100 1.32 OA-7000 1.28 OA-3200 1.28 { D, L }</p>	<p>OA-3100 1.21 { D, L } OA-3300 1.15 { L, D } OA-4800 1.05</p>	<p>AGP 1.35</p>	