

















<p>A-46</p>  <p>HSA Column:4.0mmI.D.×10cm Mobile phase: 10% 2-propanol in 100mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p>	<p>N-2,4-DNP-glutamic acid</p> 	<p>A-47</p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 1% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate: Detector:UV210nm</p>	<p>N-2,4-DNP-methionine</p> 
<p>HSA 2.83</p>		<p>AGP 2.22 HSA 2.00</p>	
<p>A-48</p>  <p>HSA Column:4.0mmI.D.×10cm Mobile phase: 1% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.5mL/min Detector:UV210nm</p>	<p>N-2,4-DNP-norleucine</p> 	<p>A-49</p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 6% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p>	<p>Etodolac</p> 
<p>HSA 7.73</p>		<p>AGP 4.73</p>	
<p>A-50</p>  <p>HSA Column:4.0mmI.D.×10cm Mobile phase: 6% 2-propanol in 100mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p>	<p>Folinic acid (Leucovorin)</p> 	<p>A-51</p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 5% acetonitrile in 10mmol/L sod.ph.b. pH5.0 Flow rate:0.5mL/min Detector:UV225nm</p>	<p>Hippurylphenyllactic acid</p> 
<p>HSA 8.32</p>		<p>AGP 1.57</p>	
<p>A-53</p>  <p>HSA Column:4.0mmI.D.×10cm Mobile phase: 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p>	<p>Kynurenine</p> 	<p>A-54</p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p>	<p>Luciferin</p> 
<p>HSA 8.62</p>		<p>AGP 3.76</p>	