

















<p>B-41 <u>Carbuterol</u></p>  <p><u>CBH</u> Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ph.b. pH7.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV210nm</p>		<p>B-42 <u>cis-trans-Cavinton</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×5cm Mobile phase: 35% acetonitrile in 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p>	
<p>CBH 1.66</p>		<p>AGP 8.80</p>	
<p>B-43 <u>Clenbuterol</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 1% 2-propanol in 10mmol/L sod.ac.b. pH5.0 (total acetate conc. = 15mmol/L) Flow rate:0.9mL/min Detector:UV225nm</p>		<p>B-44 <u>Dixyrazine</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 1% acetonitrile in 10mmol/L sod.ac.b. pH4.0 (total acetate conc. = 60mmol/L) Flow rate:0.9mL/min Detector:UV225nm</p>	
<p>AGP 1.60</p>		<p>AGP 1.41</p>	
<p>B-45 <u>Dropropizine</u></p>  <p><u>CBH</u> Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ph.b. pH7.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV210nm</p>		<p>B-46 <u>Epanolol</u></p>  <p><u>CBH</u> Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ac.b. pH5.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV210nm</p>	
<p>CBH 1.58</p>		<p>CBH 2.20</p>	
<p>B-47 <u>Ephedrine</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase:1mmol/L octanoic acid in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p>		<p>B-48 <u>E-10-hydroxy Nortriptyline</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 7.5% 2-propanol in 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p>	
<p>AGP 1.26</p>		<p>AGP 2.36</p>	