











<p>B-66</p> <p><u>Tolamolol</u></p>  <p><u>CBH</u> Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ac.b. pH5.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV225nm</p>		<p>B-67</p> <p><u>Toliprolol</u></p>  <p><u>CBH</u> Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ph.b. pH6.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV225nm</p>	
<p>CBH 2.05</p>		<p>CBH 6.69</p>	
<p>B-68</p> <p><u>Trihexyphenidyl</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 3% acetonitrile in 10mmol/L sod.ac.b. pH4.1 (total acetate conc. = 50mmol/L) Flow rate:0.9mL/min Detector:UV210nm</p>		<p>B-69</p> <p><u>Oxamniquine</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 0.6% acetonitrile in 10mmol/L sod.ph.b. pH5.2 Flow rate:0.9mL/min Detector:UV246nm</p>	
<p>AGP 1.30</p>		<p>AGP 1.95</p>	
<p>B-70</p> <p><u>Tropicamide</u></p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 3% acetonitrile in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p>			
<p>AGP 1.31</p>			