

















<p>C-08 Ketamine</p>  <p>OA-5500 Column:4.6mmI.D.×15cm Mobile phase:2mmol/L copper (II) sulfate in water/acetonitrile (85:15) Flow rate:1.0mL/min Detector:UV254nm</p> 		<p>C-09 Methoxyphenamine</p>  <p>OA-4400 Column:4.6mmI.D.×25cm Mobile phase:hexane/tetrahydrofuran/methanol/trifluoroacetic acid (60:35:5:0.2) Flow rate:1.0mL/min Detector:UV254nm Injection:2µL (1mg/mL)</p> 	
<p>OA-5500 1.26 OA-4900 1.12 AGP 1.90</p>		<p>OA-4400 1.07</p>	
<p>C-11 Carbinoxamine</p>  <p>OA-4500 Column:4.6mmI.D.×25cm Mobile phase:hexane/tetrahydrofuran/methanol/trifluoroacetic acid (60:30:10:0.2) Flow rate:1.0mL/min Detector:UV254nm Injection:2µL (1mg/mL)</p> 		<p>C-12 Homochlorocyclizine</p>  <p>OA-4900 Column:4.0mmI.D.×25cm Mobile phase:hexane/1,2-dichloroethane/methanol/trifluoroacetic acid (290:100:10:1) Flow rate:1.0mL/min Detector:UV254nm</p> 	
<p>OA-4500 1.10</p>		<p>OA-4900 1.08</p>	
<p>C-13 Verapamil</p>  <p>OA-4900 Column:4.6mmI.D.×25cm Mobile phase:hexane/2-propanol/methanol (80:10:10) Flow rate:1.0mL/min Detector:UV254nm Injection:5µL (1mg/mL)</p> 		<p>C-14 1,2-Diphenylethylenediamine</p>  <p>OA-5000 Column:4.6mmI.D.×15cm×2, 40°C Mobile phase:1mmol/L copper (II) acetate and 0.1mol/L ammonium acetate in water/2-propanol (85:15) (pH4.5) Flow rate:1.0mL/min Detector:UV280nm</p> 	
<p>OA-4900 1.12 OA-4100 1.09 OA-4500 1.08</p>	<p>AGP 1.31</p>	<p>OA-5000 1.10</p>	
<p>C-15 Tetramisole</p>  <p>OA-4600 Column:4.0mmI.D.×25cm Mobile phase:hexane/1,2-dichloroethane/methanol/trifluoroacetic acid (245:140:15:1) Flow rate:1.0mL/min Detector:UV254nm Injection:1µL (10mg/mL)</p> 		<p>C-16 PTH-valine</p>  <p>OA-4700 Column:4.6mmI.D.×25cm Mobile phase:hexane/2-propanol/methanol/trifluoroacetic acid (90:8:2:0.2) Flow rate:1.0mL/min Detector:UV254nm Injection:1µL (1mg/mL)</p> 	
<p>OA-4600 1.18 { D, L } OA-4000 1.15 OA-4100 1.09</p>	<p>OA-4700 1.09 AGP 1.20 CBH 1.95</p>	<p>OA-4700 1.24 OA-4100 1.15</p>	