

















<p><b>C-42</b>     <b>Felodipine</b></p>  <p>AGP Column:4.0mmI.D.×15cm Mobile phase: 12% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p> 	<p><b>C-43</b>     <b>Fendiline</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 3% acetonitrile in 10mmol/L sod.ac.b. pH4.1 (total acetate conc. = 60mmol/L) Flow rate:0.9mL/min     Detector:UV225nm</p> 
<p>AGP     1.33</p>	<p>AGP     1.41</p>
<p><b>C-44</b>     <b>Isradipine</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 15% 2-propanol in 30mmol/L sod.ph.b. pH6.8 Flow rate:0.9mL/min Detector:UV225nm</p> 	<p><b>C-45</b>     <b>Laudanosine</b></p>  <p>CBH Column:4.0mmI.D.×10cm Mobile phase:5% 2-propanol in 10mmol/L sod.ph.b. pH6.0 + 50µmol/L EDTA Flow rate:0.9mL/min Detector:UV225nm</p> 
<p>AGP     2.07</p>	<p>CBH     2.06</p>
<p><b>C-46</b>     <b>Medetomidine</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 1% 2-propanol in 20mmol/L sod.ac.b. pH5.0 Flow rate:0.9mL/min Detector:UV220nm</p> 	<p><b>C-47</b>     <b>Mephencytollin</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 0.3% 1-propanol in 100mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p> 
<p>AGP     1.73 { -, + }</p>	<p>AGP     1.23 HSA     1.36</p>
<p><b>C-48</b>     <b>Mepivacaine</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 9% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV225nm</p> 	<p><b>C-49</b>     <b>Methadone</b></p>  <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 10% 2-propanol in 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p> 
<p>AGP     1.31</p>	<p>AGP     1.47</p>