











D-23	<u>α-Bromophenylacetic acid</u>	D-24	<u>trans-2-Phenyl-1-cyclopropane-carboxylic acid</u>
 <p>OA-2500 Column:4.6mmI.D.×25cm Mobile phase:0.01mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p>		 <p>OA-2500 Column:4.6mmI.D. × 25cm Mobile phase:0.01mol/L ammonium acetate in methanol Flow rate:1.0mL/min Detector:UV254nm</p>	
OA-2500 1.11		OA-2500 1.08	
D-25	<u>Berabrost sodium</u>	D-26	<u>2-(2,4-Dichlorophenoxy)propionic acid</u>
 <p>AGP 4.0mmI.D.×10cm Mobile phase: 0.5% acetonitrile in 20mmol/L sod.ph.b. pH7.0 Flow rate:1.0mL/min Detector:FL EX 282nm, EM 304nm</p>		 <p>AGP Column:4.0mmI.D. × 10cm Mobile phase: 4% 2-propanol in 10mmol/L sod.ph.b. pH5.2 Flow rate:0.9mL/min Detector:UV220nm</p>	
AGP		AGP 1.26	
D-27	<u>Tiaprofenic acid</u>		
 <p>AGP Column:4.0mmI.D.×10cm Mobile phase: 1% 1-propanol in 10mmol/L sod.ph.b. pH6.5 Flow rate:0.9mL/min Detector:UV225nm</p>			
AGP 1.56			