













<p>E-09 Cyclandelate</p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 15% 2-propanol in 10mmol/L sod.ph.b. pH6.0 Flow rate:0.9mL/min Detector:UV225nm</p> 		<p>E-10 1,2,3,4-Tetrahydro-1-naphthol</p>  <p><u>AGP</u> Column:4.0mmI.D.×10cm Mobile phase: 3% 2-propanol in 10mmol/L sod.ph.b. pH7.0 Flow rate:0.9mL/min Detector:UV210nm</p> 	
<p>AGP 1.54</p>		<p>AGP 1.75</p>	
<p>E-11 Estradiol</p>  <p><u>OA-7500</u> Column:4.6mmI.D.×25cm Mobile phase: water/acetonitrile (55:45) Flow rate:0.8mL/min Detector:UV280nm</p> 		<p>E-12 Estrone</p>  <p><u>OA-7500</u> Column:4.6mmI.D.×25cm Mobile phase: water/acetonitrile (55:45) Flow rate:0.8mL/min Detector:UV280nm</p> 	
<p>OA-7500 1.14</p>		<p>OA-7500 1.10</p>	
<p>E-13 Ethinylestradiol</p>  <p><u>OA-7500</u> Column:4.6mmI.D.×25cm Mobile phase: 70% methanol in 0.1% TEAA pH4.0 Flow rate:0.7mL/min Detector:UV278nm</p> 		<p>E-14 Norgestrel</p>  <p><u>OA-7500</u> Column:4.6mmI.D.×25cm Mobile phase: 60% methanol in 0.1% TEAA pH4.0 Flow rate:0.6mL/min Detector:UV220nm</p> 	
<p>OA-7500 1.47</p>		<p>OA-7500 1.16</p>	